

420-0004-042		Sht. 1 of 1	APP'D BY SGA
ISSUE	EDO NO.	APP'D	DATE
3	11-01-246	SGA	12/14/01
4	5-02-211	SGA	6-6-02

Declaration of Conformity

We

**AMETEK DREXELBROOK
205 KEITH VALLEY ROAD
HORSHAM, PENNSYLVANIA
USA 19044**

declare under our sole responsibility that the product Verigap Point Level Measuring System Model Number 504/404-1000 Series, to which this declaration relates is in conformity with the following standards and entitled to carry the CE Mark:

Product Type: Generic Heavy Industrial

Following the provisions of 89/336 EEC Directive,

Conforms to the emissions requirements of:


- EN 55011:1998 Class A, Conducted Emissions 150 kHz to 30 MHz
- EN 55011:1998 Class B, Radiated Emissions, 30MHz to 1GHz

Conforms to the immunity requirements of EN 61000-6-2:1999:

- EN 61000-4-2:1995 Electrostatic Discharge
- EN 61000-4-3:1997 Radiated Immunity
- ENV 50204:1994 Radiated Immunity, Pulsed
- EN 61000-4-4:1995 EFT/Burst, Power and I/O Leads
- EN 61000-4-5:1995 Surge Immunity, Power Leads
- EN 61000-4-6:1996 Conducted Immunity, Power and I/O Leads
- EN 61000-4-8:1994 Power Frequency Magnetic Fields (10 A(rms)/meter)
- EN 61000-4-11:1994 Voltage Dips and Interrupts

Following the provisions of 72/23 EEC Directive,

- EN 61010-1:2001 Safety Requirements For Electrical Equipment For Measurement, Control, & Laboratory Use



Steven G. Arrfold
Quality Assurance & Product Safety Manager
Issue Date 4/01/2002